

Issued to:

QEP Australasia Pty Ltd
68 Prosperity Way
Dandenong South VIC 3175

Novitatech Test Laboratory
NATA Accreditation Number 2953

Certificate of Compliance

Product Details: Q-Assist Shower Chair with Backrest, Arms and Suction Feet (QA35)

The product listed above complies with the requirements of AS/NZS 3973:2009 (R2016) for user weight to 150kgs.

Yours sincerely



Greg Paini
Senior Test Technician
Authorised Signatory
testing@novita.org.au

Date of Issue 17/09/2024

